



Govt. of Karnataka

## BANGALORE MEDICAL COLLEGE & RESEARCH INSTITUTE

[An Autonomous Institute of Govt. of Karnataka]

Fort, K.R. Road, Bangalore – 560002.

Telephone:26700810;Tele fax 26704342

email: director\_bmcric@yahoo.co. in

No. BMCRI/Sup(1)/112/22-23



Office of the Director cum Dean  
Bangalore, Dated: 06/02/2023

### NOTIFICATION

**Sub: Inviting quotation for Supply of Video Laryngoscope to BMCRI-reg.**

Sealed quotations are invited to quote the competitive rates for the Supply of Video Laryngoscope to BMCRI before 29/02/2023.

Sl. No	Particulars	Qty	Period	Place of Delivery
01	Video Laryngoscope (Specifications enclosed)	01	One Week	Dept of Anaesthesiology, BMCRI, Bangalore

**1. Quoted price**

- The tenderer may quote for the item in the format of quotation attached;
- All duties, taxes and other levies payable by the tenderer (including Sales tax on the finished goods) shall be included in the item rate.
- The rates quoted for the item shall be fixed for the duration of the contract and shall not be subject to any adjustment.
- Rates for supply of partial quantity of the item is not acceptable.
- Corrections if any shall be made by crossing out, initialing, dating and rewriting.
- Cable or Facsimile quotations are not acceptable.

**2. a) Each Supplier must submit only one quotation**

- Suppliers should have registered with appropriate authority.
- Should have register number for GST.
- Submit Xerox copy of GST Registration copy attested along with quotation

**3. Validity of quotations:**

The quotation shall remain valid for a period not less than 30 days after the deadline fixed for submission of quotations.

**4. Evaluation of quotations:**

The Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e., which are properly signed, and conform to the terms and conditions and specifications in the following manner:

- The evaluation will be done including the Sales tax. If the tenderer has not included the Sales tax in his quotation for the item rate, and has also not indicated the rate of Sales tax applicable, the quoted rate will treated as though it is inclusive of the Sales tax and no extra payment for Sales tax will be made;
- The evaluation would be done for the item separately The tenderer who has quoted for partial quantity of the item would be treated as non-responsive. Purchaser will award the contract for the item separately to the lowest responsive tenderer for the item.

.....2/-

**5. Award of contract:**

- (a) The Purchaser will award the contract to the tenderer whose quotation has been determined to be substantially responsive and who has offered the lowest price as per Para 5(b) below.
- (b) The Purchaser reserves the right at the time of contract award to increase or decrease the quantities of the item indicated in para 1 above by 25% without any change in the unit price or any other terms and conditions.
- (c) The Purchaser prior to the expiration of the quotation validity period will notify the tenderer whose quotation is accepted of the award of contract. The terms of the accepted offer shall be incorporated in the Purchase order.
- (d) Normal commercial warranty/guarantee shall be applicable to the supplied goods;
- (e) Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the quotation process and reject all quotations at any time prior to the award of the contract.

**6. Last date and time of receipt of quotations:**

You are requested to submit the sealed quotations superscribed on the envelope as "quotation for supply of Video Laryngoscope to BMCRI Due on 20/02/2023" before 5:30PM. Wlook forward to receiving your quotations.



Director cum Dean  
Bangalore Medical College and  
Research Institute, Bangalore.

**Copy to:**

- 1 PA to Director cum Dean, BMCRI requesting to publish in website.
- 2 Medical Superintendent, Victoria Hospital / Vanivilas Hospital / Bowring & Lady Curzon Hospital / Minto Hospital / SDS TB & Rajiv Gandhi Chest Diseases Hospital /K.C.General Hospital / Jayanagar General Hospital, Bangalore with request to display this notification on notice board of your institute.
- 3 Notice Board.
- 4 To all the Registered Vendors through E-mail (Prtg. &Stsnry).

## SPECIFICATION VIDEO LARYNGOSCOPE

Display	8 Inch LCD Monitor HD Display	
Resolution	1024*768RGB	
Data audiovisual output	Support	
Camera Resolution	2M Pixel	
View Angle	66°	
Illumination	≥ 800LUX	
Battery Life Time	120-150min	
Battery Capacity	6400mAh	
Voltage	3.7V	
Blade type	Reusable Blades	
Material of Blade	SUS316L	
Shot and Record Function	Support	
Anti-fog	Anti-fog when it's on	
Blade Sizing	Miller 00/0/1; Mac 1/2/3/4/5 MacD 1/2/3/4/5	Mac 1/2/3/4/5
HDMI&USB Port	Include	
Connector for Screen and Handle	Include	
Bracket (with tray)	Include	
Attachments	Adapter, Power cord, Data connection line	
	Should be easily disinfected and reusable	

Specification video laryngoscope

<b>Display</b>	Size: 3.5inch Full view
	Monitor: LCD/ LED
	Resolution Ratio: 640X480RGB
	Aspect Ratio: 4:3
	Video refresh Rata $\geq$ 30FPS
<b>Data output</b>	Illumination: LED
	Easy to establish and store files
<b>Image output</b>	USB interface/ LED screen
<b>Laryngoscope blade Camera Include</b>	Resolution Ratio $\geq$ 2.0M Pixels
	Field Angle: 66°
	Illuminance $\geq$ 800LUX
<b>Handle</b>	Confortable ergonomic short handle design
<b>Battery</b>	Type Rechargeable Lithium Battery
	Long working time $\geq$ 120 Minutes
	Capacity > or = to 3200mAh
	Battery life Cycle: More Than 300 times
	Charging times: less than < 6 hrs